



OCT 3 - 2005

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Ms. Linda Nelson  
Regulatory Affairs Manager  
Steritec Products, Incorporated  
599 Topeka Way, Suite 400  
Castle Rock, Colorado 80109

Re: K052516

Trade/Device Name: STERITEC CONTAINER CORD – CI 135, STERITEC  
PROCESS INDICATOR

Regulation Number: 21 CFR 880.2800

Regulation Name: Sterilization process indicator

Regulatory Class: II

Product Code: JOJ

Dated: August 20, 2005

Received: September 14, 2005

Dear Ms. Nelson:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050. This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Chiu Lin", with a stylized flourish at the end.

Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

## Indications for Use

510(k) Number (if know): K052516

Device Names: SteriTec Container Card – CI 135, SteriTec Process Indicator Labels – CI 140, and SteriTec Process Indicator DOTS – CI 137.

### Indications For Use:

SteriTec Process Indicator Container Cards, Labels, and DOTS are intended for use by health care providers with steam sterilization wraps, container, cassettes, or pouches to distinguish between processed and unprocessed items. The Cards, Labels, and DOTS can be used in 121° C gravity steam sterilizers or 132° – 134° C pre-vacuum steam sterilizers. Indicators turn to dark gray/black after exposure to steam sterilization processes.

Prescription Use \_\_\_\_\_  
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use X  
(21 CFR 801 Subpart C)

**(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE  
IF NEEDED.)**

Concurrence of CDRH, Office of Device Evaluation (ODE)

[Signature]  
(Division Sign-Off)  
Division of Anesthesiology, General Hospital,  
Infection Control, Dental Devices

510(k) Number: K052516